



Mougins, France, 5 April 2019

Dear Partners,

Based on a precautionary principle, the French National Security Agency for Medicines and Health Products (ANSM) took yesterday the isolated decision to stop the sales in France and to recall on the French territory Polyurethane and Macro-Textured breast implants, impacting the 6 brands offering those products on the French market.

For our Laboratories, this is limited to our Silicone Monobloc® Textured implants (-**T** in the end of the product reference).

As you know, the first priority of the Laboratoires Arion is patient safety through product excellence. Our products impacted by this decision have been produced in total compliance with the quality and regulatory standards. This decision does not sanction any defect on the product or any failure in the production process.

This decision is limited to France only, and does not concern our Monobloc® Micro-Textured implants (-**MT** at the end of the product reference) or Monobloc® Smooth implants (-**L** at the end of the product reference). It does not concern our Monobloc® Hydrogel product range either.

We stay at your disposal for any further information,

Best regards,

Jean-Claude Arion

CEO

A handwritten signature in blue ink, consisting of several vertical strokes and a horizontal line, representing the name Jean-Claude Arion.